





# **EU Type Examination Certificate**

This is to certify that: Keeo Life Pvt Ltd

M-6 OFFICE, BALRAM HOUSE KARAMPURA

Delhi 110015 India

**Holds Certificate Number:** CE 734978

In respect of:

Model KF-SG-N95-EL and KF-SG-N95-HL Particulate filtering half masks To technical specification to Annex II (EHSR) of the PPE Regulation (EU) 2016/425 PPE for use by healthcare professionals as per Commission recommendation 2020/403

on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Drs. Dave Hagenaars, Managing Director

First Issued: 2020-10-27 Latest Issue: 2020-10-27 Effective Date: 2020-10-27 Expiry Date: 2021-10-27

Page: 1 of 3



...making excellence a habit.™

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request. To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated online.

## **EU Type Examination Certificate**

No. CE 734978

### **Product Specification**

**Product Name:** Particulate Respirator.

**Product Type:** Particulate filtering half masks for use by Healthcare professionals.

Model: KF-SG-N95-EL and KF-SG-N95-HL

Classification: KF-SG-N95-EL FFP2 NR un-valved and KF-SG-N95-HL FFP2 NR un-valved

**Technical Specification:** Technical specification to satisfy Annex II of the PPE Regulation (EU) 2016/425.

**Product Description:** The KF-SG-N95-EL respirator is non-reusable, secured to the face of the

user by a pair of elastic ear straps, and has no exhalation valve. The KF-SG-N95-HL respirator is non-reusable, secured to the face of the user by a pair of elastic head straps, and has no exhalation valve The respirators are

FFP2 class, fold flat type.

The respirator listed on this certificate is for use by healthcare workers, first

responders and other personnel involved in the efforts to contain the COVID-19 virus

and avoid its further spread.

The product covered by this certificate is not approved for industrial applications and

the certificate is only valid as long as EU Commission recommendation sheet

2020/403 remains applicable.

**Product Assessments:** BSI's PPE for Healthcare Professionals 2020/403 – RPE Technical Specification.

First Issued: 2020-10-27 Effective Date: 2020-10-27 Latest Issue: 2020-10-27 Expiry Date: 2021-10-27

Page: 2 of 3

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.

To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated online.

# **EU Type Examination Certificate**

No. CE 734978

#### **Certificate Administration Details**

Technical File Reference: KEO-PPE-01 Rev 01

#### **Certificate Amendment Record:**

| Issue date   | Comments             | <b>BSI Review No.</b> |
|--------------|----------------------|-----------------------|
| October 2020 | First issue, rev. 01 | 2797:20:3277773       |

### **Certificate validity**

The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspect of the overall process utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

The validity of the Certificate for the products is also dependent on the maintenance of the EU Conformity to Type based on Internal Production Control plus supervised product checks at random intervals, Annex VII (Module C2) as referenced on BSI issued Certificate CE 734979.

First Issued: 2020-10-27

Latest Issue: 2020-10-27

Effective Date: 2020-10-27

Expiry Date: 2021-10-27

Page: 3 of 3

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.

To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated online.